



The Tech Works **CC2-EPS** is an Emergency Pull Station in a semi-flush ABS plastic enclosure for durable emergency signaling. A Hygenicord pull cord provides the patient with call-for-help access when they may not be able to reach the station. A convenient Push for Help button is also included for activating the Call directly at the station. Staff is provided with a distinctive "CANCEL" button to reset the station.

An integral call confirmation LED notifies the user that a call for help has been sent. The Call Assurance Light will be Off if no call is pending, glowing steadily if a Normal Call is pending, and flashing if an Emergency Call is Pending.

The **CC2-EPS** is a passive button-and-light assembly designed for use with a wide range of active stations. **CC2-Series** devices operating on 11-25VDC making them compatible with the **CC2-DL-44-B** Dome Light or the **CC2-CM-4** Control Module as well as most of our legacy products. Screw Terminals provide wiring connection to the active electronic stations that provide signaling and control for the **CC2-EPS**.

Mounting is a standard 1-gang electrical box. An integral back plate provides mounting and adaptor alignment for the **CC2-EPS**.

# CC2-EPS

## Emergency Pull Station

### BENEFITS

- Reliable Signaling
- No Maintenance
- Low Power
- Easy to Operate
- Durable Construction
- Simple Wiring
- Attractive Station Hardware

### Associated Equipment

<b>CC2-DL-44-B</b>	Dome Light
<b>CC2-CM-4</b>	Control Module

### Design Information

<b>Power</b>	11-25V DC
<b>Color</b>	White
<b>Mounting</b>	1 Gang Back Box
<b>Dimensions</b>	3.25" W x 0.5" D x 5" H
<b>Weight</b>	0.25 lbs

### Architects' and Engineers' Specifications

The System Emergency Pull Station shall be a standard one gang electrical box mounting device constructed of ABS plastic. A large push for Help button shall be included with a durable nylon cord to provide both push and pull for help operation. A distinctive Cancel button shall be provided to reset the station. A call confirmation light shall be included to indicate that a call has been placed. The pull station shall be a passive electronic device requiring no more than 30 mA at 11-25 Volts DC for full operation.

The System Emergency Pull Station shall be Tech Works Model CC2-EPS.



**Tech Works**®

"Making Specialized Communication Easy"



## Quick Start Installation Instructions

