NC-SPS

Nurse Call
Staff Presence Station

The Tech Works **NC-SPS** is a Staff Presence Station in a semi-flush ABS plastic enclosure for durable emergency signaling. A "STAFF" button generates a light signal to notify staff that someone is in a room. A convenient Push for Help button is also included for activating the Emergency Call directly at the station. Staff is provided with a distinctive "CANCEL" button to reset the station.

An integral call confirmation LED notifies the user that a call has been sent. The Call Assurance Light will be Off if no call is pending, glowing steadily if a Normal Call is pending, and flashing if an Emergency Call is pending.

This is a passive button and light assembly designed for use with active stations like the **NC-DL-12-R** Dome Light. Screw Terminals provide wiring connection to the active electronic stations that provide signaling and control for the **NC-SPS**.

Mounting is a standard 1-gang electrical box. An integral back plate provides mounting and adaptor alignment for the **NC-SPS**.

BENEFITS
- Reliable Signaling
- No Maintenance
- Low Power
- Simple Wiring
- Easily Integrates to other Systems
- Attractive Station Hardware

Associated Equipment
> **NC-DL-12** Dome Light

Design Information
Power	15V DC
Color	White
Mounting	1 Gang Back Box
Dimensions	3.25" W x 0.5" D x 5" H
Weight	0.25 lbs.

Architects' and Engineers' Specifications
The Nurse Call System Staff Presence Station shall be a standard one gang electrical box mounting device constructed of ABS plastic. A large STAFF push button will let the system know that Staff is in the room. A separate push for Help button shall be included to place a Staff Needs Assistance Call to the System. A distinctive Cancel button shall be provided to reset the station. A call confirmation light shall be included to indicate that a call has been placed. The staff station shall be a passive electronic device requiring no more than 3 mA at 15 Volts DC for full operation.

The Nurse Call System Staff Presence Station shall be Tech Works Model NC-SPS

Tech Works ®

"Making Specialized Communication Easy"

Quick Start Installation Instructions

Tech Works
NC-DL-12

FLASH CONTROL
1 2 3 TONE

ON

OFF

EIA 56RR

BUSS FROM
PREVIOUS DOME LIGHT
OR
CONTROL MODULE

BUSS IN

24 VDC (+)

COMMON

FLASH / MASTER

SIGNAL

BUSS OUT

BUSS TO NEXT
DOME LIGHT

SIGNAL OUT TO
CONTROL MODULE

TONE (ON/OFF)	SLAVE & MASTER OUTPUT	STAND ALONE & MASTER OUTPUT	CONTROL MODULE & MASTER OUTPUT	CONTROL MODULE & FLASH OUTPUT	CONFIGURATION
ON/OFF	ON/OFF	ON/OFF	ON/OFF	ON/OFF	
OFF	ON	ON	OFF	OFF	SW-4
ON	OFF	ON	OFF	OFF	SW-3
ON	OFF	OFF	OFF	OFF	SW-2
ON	OFF	OFF	OFF	OFF	SW-1

STATION 1

CANCEL SW
E-CALL SW
N-CALL SW
COMMON

STATION 2

Tech Works
NC-PBS

COMMON N-CALL E-CALL CANCEL

Tech Works
NC-EPS
NC-SPS

COMMON N-CALL E-CALL CANCEL