



The Tech Works CC2-PBS is a Patient Bed Station in a semi-flush ABS plastic enclosure for durable emergency signaling. A 1/4" phone jack provides a supervised connection for a Call Cord like the PBC-7. The Call Cord provides the patient with call-for-help access when they may not be able to reach the station. A convenient Push for Help button is also included for activating the Emergency Call directly at the station. Staff is provided with a distinctive "CANCEL" button to reset the station.

An integral call confirmation LED notifies the user that a call for help has been sent. The Call Assurance Light will be Off if no call is pending. glowing steadily if a Normal Call is pending, and flashing if an Emergency Call is Pending.

The CC2-PBS is a passive button-and-light assembly designed for use with a wide range of active stations. **CC2-Series** devices 11-25VDC making operating on compatible with the CC2-DL-44-B Dome Light or the CC2-CM-4 Control Module as well as most of our legacy products. Screw Terminals provide wiring connection to the active electronic stations that provide signaling and control for the CC2-PBS.

Mounting is a standard 1-gang electrical box. An integral back plate provides mounting and adaptor alignment for the CC2-PBS.

CC2-PBS

Patient Bed Station

BENEFITS

- Reliable Signaling
- No Maintenance
- Low Power
- Easy to Operate
- **Durable Construction**
- Simple Wiring
- Attractive Station Hardware

Associated Equipment

PBC-7 Patient Call Cord **PBC-2-12 Dual Patient Call**

Cord

CC2-DL-44-B Dome Light CC2-CM-4 Control Module

Design Information

Power 11-25V DC Color White

Mounting 1 Gang Back Box 3.25" W x 0.5" D x 5" H **Dimensions**

Weight 0.25 lbs.

Architects' and Engineers' Specifications

The System Patient Bed Station shall be a standard one gang electrical box mounting device constructed of ABS plastic. A large push for Help button shall be included with a 1/4: phone Jack for a call cord to provide both local and remote push for help operation. A distinctive Cancel button shall be provided to reset the station. A call confirmation light shall be included to indicate that a call has been placed. The patient station shall be a passive electronic device requiring no more than 30 mA at 11-25 Volts DC for full operation.

The System Patient Bed Station shall be Tech Works Model CC2-PBS



Tech Works®



Quick Start Installation Instructions



